



The Open Anesthesia Journal

Content list available at: <https://openanesthesiajournal.com>



REVIEW ARTICLE

Acoustical Respiratory Monitoring in the Time Domain

D. John Doyle^{1,*}

¹Department of General Anesthesiology, Cleveland Clinic Abu Dhabi, Abu Dhabi, UAE

Abstract: This paper introduces the reader to some of the various methods that are available for the time-domain bio-acoustical monitoring of patient breathing. Technical details concerning microphone selection, calibration and characterization, signal amplification, signal filtering and waveform recording are presented. We also describe proof of concept recordings obtained from the neck, from the external ear canal, from a microphone embedded into an oxygen mask and from a leak-free microphone pneumatically connected to the cuff of a laryngeal mask airway. We recommend Audacity, an open-source digital audio editor and recording package that can be freely downloaded at <https://www.audacityteam.org> for investigators seeking to conduct research on breath sound analysis.

Keywords: Bio-acoustics, Breath sounds, Crackles, Electronic sound recording, Shannon entropy, Stridor, Wheezes, Masimo Acoustic RR.

Article History

Received: September 19, 2019

Revised: October 02, 2019

Accepted: November 09, 2019

1. INTRODUCTION

In contemporary medicine the need for respiratory monitoring has become especially important with the heavy use of opioids for perioperative pain management [1 - 3]. Despite this clinical need, no simple dependable method of continuous respiratory monitoring has come into routine clinical use, although the Masimo system of respiratory monitoring [4] and capnography (CO₂ monitoring) come close. Capnography is amongst the most popular method of continuous respiratory monitoring but suffers from a need to continually ensure that the gas sampling system is operating correctly [5]. Extraction of respiratory information from the pulse oximeter photoplethysmograph signal remains a field of active research [6, 7] but has not yet come into the mainstream. A device known as the ExSpirom 1Xi Monitor Pack (Respiratory Motion Inc., Waltham, MA) [8 - 10] operates by detecting electrical impedance changes in the thorax and respiratory muscles and seems to be particularly promising, although it is neither simple nor inexpensive.

In addition to reviewing some of the practical fundamentals of practical acoustical respiratory monitoring this report provides information of potential value to investigators who wish to develop up their own system for digital recording and analysis of breath sounds with a view to exploring how acoustical methods of respiratory monitoring might be improved. Preliminary results are provided in this report on the potential value of studying breath sounds recorded from

various locations: from the neck, from the external ear canal, from a microphone embedded into an oxygen mask and from a leak-free microphone pneumatically connected to the cuff of a laryngeal mask airway.

1.1. Breath Sounds

Breath sounds are the result of laminar and turbulent gas flow in airway structures [11, 12]. Normal breath sounds are often classified as tracheal, bronchial, bronchovesicular, and vesicular sounds, depending on where the auscultation is carried out. For example, tracheal breath sounds obtained over the trachea are sometimes described as harsh, like the sound of air is being blown through a pipe, while vesicular sounds are sometimes described as “soft, blowing, or rustling sounds” normally found throughout inspiration, and continuing, fading about one third of the way through expiration.¹ In addition to descriptions based on their location, normal breath sounds are often described by their duration, their intensity, their pitch (frequency content) and their timing (*e.g.*, inspiratory vs expiratory) within the respiratory cycle.

Pathological changes sometimes change the characteristics of lung sounds. Under such conditions, the regular lung sounds may contain additional superimposed pathological sounds, known as adventitious sounds. If they are mostly continuous, they are called “wheezing” when high pitched, and “rhonchi” when low pitched. Wheezing is often found in asthmatic patients suffering from bronchospasm. Rhonchi are often found heard in patients with Chronic Obstructive Pulmonary Disease (COPD), bronchiectasis, pneumonia, chronic bronchitis, and cystic fibrosis.

* Address correspondence to this author at Department of General Anesthesiology, Cleveland Clinic Abu Dhabi, Abu Dhabi; UAE; Tel: 052 6997627; Email: djdoyle@hotmail.com

¹ <http://www.rnceus.com/resp/respnorm.html>

If the adventitious sounds tend to be discontinuous and “explosive” in character they are often known as fine crackles (or rales, to use the old term) when high pitched, and as coarse crackles (or rales) when low pitched. Crackles are often found in a lung field that has fluid in the small airways and are thought to be due to the sudden reopening of the small airways previously closed by surface forces [11]. The characteristics of crackles sometimes offer helpful diagnostic information [13]. For example, crackles that do not clear after a cough are suggestive of pulmonary edema or pulmonary fibrosis (and several other conditions) while crackles that clear or change after coughing are more commonly found (for example) in bronchiectasisⁱⁱ.

1.2. Apnea Detection

Liu *et al.* [14] recorded tracheal sounds from 121 Post-Anesthesia Care Unit (PACU) patients using a microphone encased in a plastic bell along with a processed nasal pressure signal used as a reference method. The logarithm of the tracheal sound variance was used for apnea detection. Their new algorithm detected apneas with 92% sensitivity and 98% specificity compared to the reference method.

Other applications of tracheal sound analysis include detecting episodes of sleep apnea [15], and apneic episodes during procedural sedation [16]. MacGregor *et al.* [17] conducted a statistical analysis of tracheal breath sounds during wakefulness with a view to screening for obstructive sleep apnea.

2. THE RESPIRATORY CYCLE

The respiratory cycle is typically described as consisting of four phases that repeat endlessly (Fig. 1)

(1) **Inspiratory Phase:** From the moment that air is drawn into the lungs to the moment that inspiratory airflow stops.

(2) **Inspiratory Pause:** From the moment that inspiratory airflow stops to the beginning of the expiratory phase.

(3) **Expiratory Phase:** From the moment that air exits the lungs to the moment that expiratory airflow stops.

(4) **Expiratory Pause:** From the moment that expiratory airflow stops to the beginning of the next inspiratory phase.

One of the challenges in a number of clinical monitoring settings is to accurately identify the above respiratory phases. For example, in some studies of blood pressure variability and heart-rate variation blood pressure wave and ECG recordings are separated into inspiratory and expiratory phases to allow documentation of cardio-respiratory interactions. Although this can be achieved relatively easily using spirometry or impedance recordings, the prospect of automatically separating lung sound signals into inspiratory and expiratory components without additional instrumentation has also met with success, for example using Shannon Entropy of respiratory sound data (Fig. 2).

2.1. Using Microphones to Indirectly Listen to Breath Sounds

Both precordial stethoscopes and esophageal stethoscopes can be to directly listen to heart sounds as well as breath sounds. In addition to using purely acoustic devices, monitoring of bioacoustical phenomena can also be achieved using electronic means, incorporating a microphone and amplifier. Both wired and wireless commercial products are available, as well as units based on Bluetooth technology [18, 19]. Electronic stethoscopes are available from companies such as **Cardionics** (cardionics.com), **Ekuore** (www.ekuore.com) and **ThinkLabs** (https://www.thinklabs.com/), although some of these products focus on cardiac applications. ThinkLabs offers a library of recorded breath sounds (lung sounds) at https://www.thinklabs.com/lung-sounds that many clinicians may find to be useful.




Fig. (1). Phases of the respiratory cycle illustrated via breath sound patterns. The vertical axis is breath sound intensity, while the horizontal axis is time in seconds. Note how the intensity of the early expiratory breath sounds in this case are stronger than the intensity of the inspiratory breath sounds and tend to diminish throughout expiration. From Abushakra A, Faezipour M. Acoustic signal classification of breathing movements to virtually aid breath regulation. *IEEE J Biomed Health Inform.* 2013; 17(2): 493-500. doi: 10.1109/JBHI.2013.2244901. PubMed PMID: 24235120.

ⁱⁱ For a sampling of normal and pathological breath sounds, the interested reader is directed to <https://soundcloud.com/sam-alsmadi>, http://www.cvmb.colostate.edu/clinsci/callan/breath_sounds.htm and <https://www.thinklabs.com/lung-sounds>




Fig. (2). Shannon Entropy of the tracheal sounds acquired using a smartphone. **Top:** Segment of tracheal sound and corresponding airflow from spirometer (positive lobes are inspirations and negative lobes are expirations); **Middle:** Volume signal obtained with the spirometer as the integral of the flow; **Bottom:** Shannon entropy of tracheal sound. Observe that local minima of the Shannon entropy are obtained around the onset of each respiratory phase. Image and figure legend from Reyes BA, Reljin N, Chon KH. Tracheal sounds acquisition using smartphones. *Sensors (Basel)*. 2014;14(8):13830-50. Published 2014 Jul 30. doi:10.3390/s140813830 and used under the terms and conditions of the Creative Commons Attribution license (<http://creativecommons.org/licenses/by/3.0/>).

Investigators seeking to setup a research program involving bio-acoustical phenomena often employ a miniature electret microphone that is easily connected to stethoscope heads or other equipment. Much of the early work at our laboratory utilized a RadioShack 33-3013 miniature omnidirectional electret microphone which was secured to a stethoscope head (or other device) under test *via* a short piece of tubing and then amplified using an ordinary audio amplifier such as a RadioShack miniature audio amplifier (Catalog # 2771008). While this arrangement works well, the recent commercial availability of USB-connected miniature electret microphone has made it even easier to obtain quality bio-acoustical recordingsⁱⁱⁱ. This is because USB microphones require no built-in battery (it obtains its power instead from the USB connection) and contains a built-in amplifier as well. An additional advantage of USB type microphones is that the final signal obtained does not depend on the characteristics of the analog and digital circuitry within the host computer. One

potential disadvantage, however, is the lack of a gain adjustment control in the built-in amplifier in most USB type microphones.

2.2. Recording Sites for Breath Sounds

There exist several locations from which breath sounds can be recorded, such as at the neck (Fig. 3), over the trachea, and at the anterior or posterior chest walls. In addition to these conventional sites, we have also obtained quality breath sound recordings embedding a miniature electret microphone into an oxygen mask (Fig. 4) as well as sounds of somewhat lower quality by placing a modified electret microphone assembly into the external ear canal (Fig. 5). Finally, we have shown that it is even possible to record breath sounds from a supraglottic airway [20]. (For a sample recording of breath sounds (and other sounds) obtained from a supraglottic airway placed in a surgical patient undergoing general anesthesia visit <http://lmamonitor.homestead.com/>).



Fig. (3). Acoustic respiratory rate monitoring device from Masimo. (a) Acoustic monitoring sensor placed on the neck via an adhesive coating. (b) The display (Masimo Rad-87) showing arterial oxygen tension, heart rate and respiratory rate respectively. Image used under a Creative Commons Attribution License 4.0 from Suzuki T, Tsuda S, Nakae H, *et al.* Usefulness of Acoustic Monitoring of Respiratory Rate in Patients Undergoing Endoscopic Submucosal Dissection. *Gastroenterol Res Pract.* 2015;2016:2964581.PMID: 26858748.

ⁱⁱⁱ. For example, see <https://www.amazon.com/Microphone-ZAFFIRO-Lavalier-Recording-Interviews/dp/B0747F2HHG>



Fig. (4). A miniature electret microphone embedded into an oxygen mask can be used to record breath sounds.




Fig. (5). Modified electret microphone assembly and placement into the external ear canal.

3. MICROPHONE CALIBRATION

In research settings it can sometimes be necessary to establish the output characteristics of a microphone when it is subjected to a known acoustic signal^{iv}. For small microphones this can be most easily achieved by inserting the microphone into a calibrator such as the unit shown in Fig. (6) and measuring the microphone output using an oscilloscope or by other means. The unit in Fig. (6) provides a 94 dB \pm 0.5 dB (SPL) reference acoustic reference signal at 1000 Hz.

For larger microphones, an alternative calibration method is “free field calibration” whereby a 1000 Hz tone (for example) is fed to an amplifier/loudspeaker assembly to generate an acoustic signal filling the room. The signal level at a particular point in the room is then measured using a Sound Pressure Level meter, such as the unit shown in Fig. (7). The microphone under test is then placed beside the Sound Pressure Level meter that is measuring the intensity of the acoustic reference signal and the output of the microphone is then measured and calibrated against the reading of the Sound Pressure Level meter.



Fig. (6). This sound level calibrator from Extech (Model 407744) provides a 94 dB \pm 0.5 dB (SPL) acoustic reference signal at 1000 Hz. More information on this unit is available at http://www.extech.com/resources/407744_407766_DS-en.pdf. **Image Credit:** <http://www.extech.com/407744/>.



Fig. (7). This sound level meter from Extech (Model 407730) measures sound intensity from 40 to 130dB SPL with an accuracy of 2dB. More information on this unit is available at http://www.extech.com/resources/407730_DS-en.pdf. **Image Credit:** <http://www.extech.com/images/407730.png>.

4. ELECTRONICALLY RECORDING BREATH SOUNDS USING A HAND-HELD RECORDER

High-fidelity hand-held audio recording systems are available that are quite suitable for the digital recording of bio-acoustical signals. Typically, these support two or more audio channels and record the data in either the WAV or MP3 audio formats (*vide infra*). Automatic gain control is a common feature in such products, but this feature may need to be disabled when calibrated recordings are desired. As an example, the battery-operated TASCAM DR-40 Digital Recorder used in our laboratory offers both built-in microphones as well as support for external microphone inputs in XLR and TRS format. The unit also allows for recordings to an SD memory card to be made in both WAV and MP3 recording formats.

^{iv} For detailed technical discussion on the subject of microphone calibration visit http://www.pcb.com/Contentstore/mktgcontent/WhitePapers/WPL_36_Acoustic_methods_calibration_PCB.pdf

5. ELECTRONICALLY RECORDING BREATH SOUNDS USING A SMARTPHONE

Smartphones can be used to record breath sounds using an external microphone in two ways. The first way to connect an external microphone to the input jack of the smartphone normally used to attach a headset. The headset jack of the smart phone typically has a ground and three connections, two connections being for the left and right headphone channels and the remaining connection being a microphone input. To use this approach for a custom design it is necessary to begin by searching for information pertaining to the pinout of the headset jack of the smart phone in question as well as determine the required characteristics of any attached microphone (e.g., some microphones require a 1.5 or 2-volt power source). While this approach has promise, it can require considerable effort as the required information is not always readily available.

A second more practical approach is to purchase an external microphone specifically designed for use with a particular smartphone, such as the iPhone. Inexpensive examples from amazon.com are <https://www.amazon.com/dp/B074KJ1N3D> (single microphone unit) and <https://www.amazon.com/dp/B071X6H6ZS> (dual microphone unit).

A third approach is to use a Bluetooth microphone, but these are hard to find in the small size needed for bio-acoustics work.

As noted earlier, a final approach is to use a USB lavalier microphone, or a USB audio mixer connected to an Android smartphone *via* an OTG adapter^v. The USB approach has the important advantage that recalibration is not needed when switching from one computer or smartphone to another, as the data is digitized in the microphone unit itself rather than in the computer or smartphone.

Reyes *et al.* [21] described their experience in acquiring tracheal sounds using the Samsung Galaxy S4 and the iPhone 4s smartphones to acquire tracheal sounds from nine healthy volunteers at airflows from 0.5 to 2.5 L/s measured using a spirometer. Tracheal sounds were acquired at the neck using a miniature electret microphone encased in a plastic bell and recorded using the built-in audio recorder application of each smartphone (Voice Recorder in the Galaxy S4, and Voice Memos in the iPhone 4s). These recordings were transferred to a personal computer for conversion to .wav format and stored for further processing using the MATLAB software environment. The investigators “found that the amplitude of the smartphone-acquired sounds was highly correlated with the airflow from a spirometer” and that the increase in sound amplitude with flow followed a power law relationship. Additionally, and perhaps more importantly, the authors found that accurate respiratory rates can be obtained from tracheal sounds when the Shannon entropy of the tracheal sounds is used as an indicator of the onset of a breath (Fig. 2).

^v A USB OTG (On-The-Go) adapter allows tablets or smartphones to act as a host, allowing other USB devices, such as flash drives, mice or keyboards, to be attached to them.

6. WAV, MP3 AND OTHER AUDIO RECORDING FORMATS

Computer files containing audio signals exist in a rich variety of formats. The specific digital layout of the audio data is called the audio coding format and comes in two types: uncompressed or compressed, the latter which can dramatically reduce the file size, but often at the price of some data loss. Uncompressed audio formats include the WAV, AIFF, and AU formats, of which the WAV audio format is by far the best known. A number of lossless compressed formats are also available, the FLAC format being amongst the best known. Amongst lossy compressed audio formats, the MP3 format is easily the best known and most commonly used for music applications. The Audacity audio editing package (Fig. 8) natively supports the WAV, MP3 and AIFF formats as well as numerous other formats through available extensions. That being said, most bio-acoustics researchers will be happy to use the WAV format for all their research work, as they are the standard audio file format used in Windows PCs and allow for CD-quality sound files. Note that the theoretical drawback of large WAV file sizes (around 10 MB per minute) is no longer a major issue given the enormous drop in data storage costs in recent years. For more information of audio file formats, the interested reader is directed to https://en.wikipedia.org/wiki/Audio_file_format.

7. SOFTWARE FOR AUDIO SIGNAL PROCESSING

Investigators who are interested in processing breath sounds and other bioacoustical signals as part of their research have available to them a true cornucopia of options. In our work spanning two decades, we have used two inexpensive audio signal processing products, Goldwave (www.goldwave.com) and Spectrogram16 (no longer commercially available.) In recent years a large number of additional products have become available, although most are aimed at sound engineers and audio producers rather than bioacoustics researchers. For a list of free software for audio applications, visit https://en.wikipedia.org/wiki/Comparison_of_free_software_for_audio. However, if one were to pick one free audio editor to master, we would recommend Audacity (Fig. 8), an open-source digital audio editor and recording application package available for the Windows, macOS/OS X and Linux operating systems. Audacity can be downloaded at <https://www.audacityteam.org/>

8. ELECTRONIC PROCESSING OF BREATH SOUNDS

Numerous investigators have looked at ways of filtering breath sounds for particular clinical ends. For instance, a number have studied methods for the separation heart sounds from lung sound signals with the goal of getting better auscultation results [22 - 26]. Other investigators have applied digital techniques to characterize respiratory crackles [27 - 31] or to detect or characterize wheezes [32, 33].




Fig. (8). Audacity is a particularly well-supported and well-documented open-source digital audio editor and recording application available for the Windows, macOS/OS X and Linux operating systems. It can be freely downloaded at <https://www.audacityteam.org>. Shown here are a series of breath sounds recorded at the neck in the author's laboratory and displayed in the time-domain using Audacity. Although Audacity has an almost overwhelming array of features and options, three classes of features found in the "Effect" menu are likely to be of special interest to bio-acoustics investigators [1]: the **Amplify** feature that scales the signal [2], the **Normalize** feature that "normalizes" the signal to a chosen maximum amplitude (e.g., 0 dB), and [3] three digital filtering options (**Low Pass Filter**, **High-Pass Filter** and **Notch Filter**). One final feature (**Change Tempo**) that may captivate some users is the ability to change the speed of a recording without changing the sound pitch, a feature of potential value in listening to heart sounds in tachycardic patients as well as to shorten the duration of verbal presentations.

One of the most fundamental forms of electronic processing of breath sounds is frequency filtering. In the past, such filtering was done primarily using analog filters, but digital filtering methods are more commonly used now because their ease of use and minimal cost. In particular, the (free) Audacity audio signal processing system offers a number of digital filtering options that investigators may find to be attractive (*vide infra*).

9. THE MASIMO SYSTEM FOR ACOUSTICAL RESPIRATORY MONITORING

The Masimo Corporation (www.masimo.com) has successfully commercialized a form of respiratory acoustic monitoring that employs a flat adhesive acoustic sensor applied to the neck, near the trachea (Fig. 3) [34 - 40]. Known as Masimo Rainbow SET Acoustic Monitoring, the system reliably estimates the connected patient's respiratory rate with the assistance of special proprietary software algorithms collectively known as Signal Extraction Technology® (SET®). While multiple studies have confirmed the value of this method of acoustical respiratory monitoring in a number of settings, the technology as currently available has some important limitations that should be explained. First, neither the raw nor the processed acoustic signal is available to the clinician to listen to, although the time-domain signal is displayed. The fact that the system does not provide an analog signal output for such purposes also limits the kind of supplementary analysis that might otherwise be performed, such as digitally recording the

obtained signals, subjecting the signal to analog or digital filtering, or carrying out real-time color spectrographic analysis of the obtained breath sounds. Also, because of the proprietary nature of the Masimo acoustic monitoring system, little is publicly known about the flat acoustic sensor used. The likelihood, however, is that their sensor is based on piezo-electric film technology, as this technology has proven to be very useful in a variety of clinical applications [41 - 43].

CONCLUSION

Although a variety of methods are available for the assessment of patient breathing, bio-acoustical methods of respiratory assessment offer some special advantages. To this end we reviewed various clinical and technical matters pertaining to recording and analyzing breath sounds in the time-domain. Audacity, an open-source digital audio editor and recording package can be freely downloaded at <https://www.audacityteam.org> and is recommended for investigators seeking to conduct research on breath sound analysis.

LIST OF ABBREVIATIONS

| | | |
|------|---|------------------------------|
| ECG | = | Electrocardiogram |
| OTG | = | On The Go |
| PACU | = | Postanesthesia Care Unit |
| RR | = | Respiratory Rate |
| SET | = | Signal Extraction Technology |
| SPL | = | Sound Pressure Level |

USB = Universal Serial Bus

CONSENT FOR PUBLICATION

Not applicable.

FUNDING

None.

CONFLICT OF INTEREST

The author declares no conflict of interest, financial or otherwise.

ACKNOWLEDGEMENTS

Declared none.

REFERENCES

- [1] Doyle DJ, Vicente KJ. Electrical short circuit as a possible cause of death in patients on PCA machines: Report on an opiate overdose and a possible preventive remedy. *Anesthesiology* 2001; 94(5): 940. [http://dx.doi.org/10.1097/00000542-200105000-00050] [PMID: 11388559]
- [2] Death by PCA. *AORN J* 2014; 99(6): 832, 782. [http://dx.doi.org/10.1016/j.aorn.2014.03.010]
- [3] Yi Y, Kang S, Hwang B. Drug overdose due to malfunction of a patient-controlled analgesia machine -A case report. *Korean J Anesthesiol* 2013; 64(3): 272-5. [http://dx.doi.org/10.4097/kjae.2013.64.3.272] [PMID: 23560197]
- [4] van Loon K, Peelen LM, van de Vlasakker EC, Kalkman CJ, van Wolfswinkel L, van Zaane B. Accuracy of remote continuous respiratory rate monitoring technologies intended for low care clinical settings: A prospective observational study. *Can J Anaesth* 2018; 65(12): 1324-32. [http://dx.doi.org/10.1007/s12630-018-1214-z] [PMID: 30194672]
- [5] Gallagher JJ. Capnography monitoring during procedural sedation and analgesia. *AACN Adv Crit Care* 2018; 29(4): 405-14. [http://dx.doi.org/10.4037/aacnacc2018684] [PMID: 30523011]
- [6] Nakajima K, Tamura T, Miike H. Monitoring of heart and respiratory rates by photoplethysmography using a digital filtering technique. *Med Eng Phys* 1996; 18(5): 365-72. [http://dx.doi.org/10.1016/1350-4533(95)00066-6] [PMID: 8818134]
- [7] Cernat RA, Ciorecan SI, Ungureanu C, Arends J, Strungaru R, Ungureanu GM. Recording system and data fusion algorithm for enhancing the estimation of the respiratory rate from photoplethysmogram. *Conf Proc IEEE Eng Med Biol Soc* 2015; 2015: 5977-80. [http://dx.doi.org/10.1109/EMBC.2015.7319753] [PMID: 26737653]
- [8] Schumann R, Kwater AP, Bonney I, *et al.* Respiratory volume monitoring in an obese surgical population and the prediction of postoperative respiratory depression by the STOP-bang OSA risk score. *J Clin Anesth* 2016; 34: 295-301. [http://dx.doi.org/10.1016/j.jclinane.2016.04.029] [PMID: 27687395]
- [9] Williams GW, George CA, Harvey BC, Freeman JE. A comparison of measurements of change in respiratory status in spontaneously breathing volunteers by the exSpirom noninvasive respiratory volume monitor versus the capnostream capnometer. *Anesth Analg* 2016; 124(1): 120-6. [http://dx.doi.org/10.1213/ANE.0000000000001395] [PMID: 27384980]
- [10] Ebert TJ, Middleton AH, Makhija N. Ventilation monitoring during moderate sedation in GI patients. *J Clin Monit Comput* 2017; 31(1): 53-7. [http://dx.doi.org/10.1007/s10877-015-9809-1] [PMID: 26628270]
- [11] Bohadana A, Izbicki G, Kraman SS. Fundamentals of lung auscultation. *N Engl J Med* 2014; 370(8): 744-51. [http://dx.doi.org/10.1056/NEJMra1302901] [PMID: 24552321]
- [12] Tawhai MH, Lin C-L. Airway gas flow. *Compr Physiol* 2011; 1(3): 1135-57. [http://dx.doi.org/10.1002/cphy.c100020] [PMID: 23733638]
- [13] Piirilä P, Sovijärvi AR. Crackles: Recording, analysis and clinical significance. *Eur Respir J* 1995; 8(12): 2139-48. [http://dx.doi.org/10.1183/09031936.95.08122139] [PMID: 8666111]
- [14] Liu J, Ai C, Zhang B, *et al.* Tracheal sounds accurately detect apnea in patients recovering from anesthesia. *J Clin Monit Comput* 2018; 33(3): 437-4. [http://dx.doi.org/10.1007/s10877-018-0192-6] [PMID: 30099704]
- [15] Penzel T, Sabil A. The use of tracheal sounds for the diagnosis of sleep apnoea. *Breathe (Sheff)* 2017; 13(2): e37-45. [http://dx.doi.org/10.1183/20734735.008817] [PMID: 29184596]
- [16] Yu L, Ting C-K, Hill BE, *et al.* Using the entropy of tracheal sounds to detect apnea during sedation in healthy nonobese volunteers. *Anesthesiology* 2013; 118(6): 1341-9. [http://dx.doi.org/10.1097/ALN.0b013e318289bb30] [PMID: 23407106]
- [17] MacGregor CA, Karimi D, Azarbarzin A, Moussavi Z. Statistical analysis of tracheal breath sounds during wakefulness for screening obstructive sleep apnea. *Conf Proc IEEE Eng Med Biol Soc* 2013; 2013: 4549-52. [http://dx.doi.org/10.1109/EMBC.2013.6610559] [PMID: 24110746]
- [18] Mondal H, Mondal S, Saha K. Development of a Low-Cost wireless phonocardiograph with a bluetooth headset under resource-Limited conditions. *Med Sci (Basel)* 2018; 6(4):E117 [http://dx.doi.org/10.3390/medsci6040117] [PMID: 30563004]
- [19] Hoon Lim K, Duck Shin Y, Hi Park S, *et al.* Correlation of blood pressure and the ratio of S1 to S2 as measured by esophageal stethoscope and wireless bluetooth transmission. *Pak J Med Sci* 2013; 29(4): 1023-7. [PMID: 24353680]
- [20] John Doyle D. The Laryngeal Mask Airway audio monitor. Obtaining breath sounds from the Laryngeal Mask Airway: A new device for patient monitoring. *Anesthesiology* 2003; 99(1): 242. [http://dx.doi.org/10.1097/00000542-200307000-00051] [PMID: 12826876]
- [21] Reyes BA, Reljin N, Chon KH. Tracheal sounds acquisition using smartphones. *Sensors (Basel)* 2014; 14(8): 13830-50. [http://dx.doi.org/10.3390/s140813830] [PMID: 25196108]
- [22] Chien J-C, Huang M-C, Lin Y-D, Chong FC. A study of heart sound and lung sound separation by independent component analysis technique. *Conf Proc IEEE Eng Med Biol Soc* 2006; 1: 5708-11. [http://dx.doi.org/10.1109/IEMBS.2006.260223] [PMID: 17945913]
- [23] Pourazad M, Moussavi Z, Farahmand F, Ward R. Heart sounds separation from lung sounds using independent component analysis. *Conf Proc IEEE Eng Med Biol Soc* 2005; 3: 2736-9. [http://dx.doi.org/10.1109/IEMBS.2005.1617037] [PMID: 17282806]
- [24] Nersissson R, Noel MM. Heart sound and lung sound separation algorithms: A review. *J Med Eng Technol* 2017; 41(1): 13-21. [http://dx.doi.org/10.1080/03091902.2016.1209589] [PMID: 27420021]
- [25] Falk TH, Chan W-Y. Modulation filtering for heart and lung sound separation from breath sound recordings. *Conf Proc IEEE Eng Med Biol Soc* 2008; 2008: 1859-62. [http://dx.doi.org/10.1109/IEMBS.2008.4649547] [PMID: 19163050]
- [26] Pourazad MT, Moussavi Z, Thomas G. Heart sound cancellation from lung sound recordings using time-frequency filtering. *Med Biol Eng Comput* 2006; 44(3): 216-25. [http://dx.doi.org/10.1007/s11517-006-0030-8] [PMID: 16937163]
- [27] Speranza CG, Moraes R. Instantaneous frequency based index to characterize respiratory crackles. *Comput Biol Med* 2018; 102: 21-9. [http://dx.doi.org/10.1016/j.compbiomed.2018.09.007] [PMID: 30240835]
- [28] Becker KW, Scheffer C, Blanckenberg MM, Diacon AH. Analysis of adventitious lung sounds originating from pulmonary tuberculosis. *Conf Proc IEEE Eng Med Biol Soc* 2013; 2013: 4334-7. [http://dx.doi.org/10.1109/EMBC.2013.6610505] [PMID: 24110692]
- [29] Abbas A, Fahim A. An automated computerized auscultation and diagnostic system for pulmonary diseases. *J Med Syst* 2010; 34(6): 1149-55. [http://dx.doi.org/10.1007/s10916-009-9334-1] [PMID: 20703592]
- [30] Dorantes-Méndez G, Charlestone-Villalobos S, González-Camarena R, Chi-Lem G, Carrillo JG, Aljama-Corralles T. Crackles detection using a time-variant autoregressive model. *Conf Proc IEEE Eng Med Biol Soc* 2008; 2008: 1894-7. [http://dx.doi.org/10.1109/IEMBS.2008.4649556] [PMID: 19163059]
- [31] Mendes L, Vogiatzis IM, Perantoni E, *et al.* Detection of crackle events using a multi-feature approach. *Conf Proc IEEE Eng Med Biol Soc* 2016; 2016: 3679-83. [http://dx.doi.org/10.1109/EMBC.2016.7591526] [PMID: 28269092]
- [32] Puder LC, Wilitzki S, Bühner C, Fischer HS, Schmalisch G.

- Computerized wheeze detection in young infants: Comparison of signals from tracheal and chest wall sensors. *Physiol Meas* 2016; 37(12): 2170-80. [http://dx.doi.org/10.1088/0967-3334/37/12/2170] [PMID: 27869106]
- [33] Bokov P, Mahut B, Flaud P, Delclaux C. Wheezing recognition algorithm using recordings of respiratory sounds at the mouth in a pediatric population. *Comput Biol Med* 2016; 70: 40-50. [http://dx.doi.org/10.1016/j.compbiomed.2016.01.002] [PMID: 26802543]
- [34] Eisenberg ME, Givony D, Levin R. Acoustic respiration rate and pulse oximetry-derived respiration rate: A clinical comparison study. *J Clin Monit Comput* 2018; 1-8. [http://dx.doi.org/10.1007/s10877-018-0222-4] [PMID: 30478523]
- [35] Kawanishi H, Inoue S, Kawaguchi M. A retrospective analysis of oxygen desaturation during acoustic respiratory rate monitoring in Non-ICU patients following tracheal extubation after general anesthesia. *Anesthesiol Res Pract* 2017; 20174203156 [http://dx.doi.org/10.1155/2017/4203156] [PMID: 28487734]
- [36] Patino M, Kalin M, Griffin A, *et al.* Comparison of postoperative respiratory monitoring by acoustic and transthoracic impedance technologies in pediatric patients at risk of respiratory depression. *Anesth Analg* 2017; 124(6): 1937-42. [http://dx.doi.org/10.1213/ANE.0000000000002062] [PMID: 28448390]
- [37] Yang S, Menne A, Hu P, *et al.* Acoustic sensor *versus* electrocardiographically derived respiratory rate in unstable trauma patients. *J Clin Monit Comput* 2017; 31(4): 765-72. [http://dx.doi.org/10.1007/s10877-016-9895-8] [PMID: 27270963]
- [38] McGrath SP, Pyke J, Taenzer AH. Assessment of continuous acoustic respiratory rate monitoring as an addition to a pulse oximetry-based patient surveillance system. *J Clin Monit Comput* 2017; 31(3): 561-9. [http://dx.doi.org/10.1007/s10877-016-9884-y] [PMID: 27142098]
- [39] Suzuki T, Tsuda S, Nakae H, *et al.* Usefulness of acoustic monitoring of respiratory rate in patients undergoing endoscopic submucosal dissection. *Gastroenterol Res Pract* 2016; 20162964581 [http://dx.doi.org/10.1155/2016/2964581] [PMID: 26858748]
- [40] Ouchi K, Fujiwara S, Sugiyama K. Acoustic method respiratory rate monitoring is useful in patients under intravenous anesthesia. *J Clin Monit Comput* 2017; 31(1): 59-65. [http://dx.doi.org/10.1007/s10877-015-9822-4] [PMID: 26759335]
- [41] Lang C, Fang J, Shao H, Ding X, Lin T. High-sensitivity acoustic sensors from nanofibre webs. *Nat Commun* 2016; 7: 11108. [http://dx.doi.org/10.1038/ncomms11108] [PMID: 27005010]
- [42] Ringgaard E, Lautzenhiser F, Bierregaard LM, Zawada T, Molz E. Development of porous piezoceramics for medical and sensor applications. *Materials (Basel)* 2015; 8(12): 8877-89. [http://dx.doi.org/10.3390/ma8125498] [PMID: 28793753]
- [43] Park J-H, Jang D-G, Park JW, Youm S-K. Wearable sensing of In-Ear pressure for heart rate monitoring with a piezoelectric sensor. *Sensors (Basel)* 2015; 15(9): 23402-17. [http://dx.doi.org/10.3390/s150923402] [PMID: 26389912]

© 2019 D. John Doyle.

This is an open access article distributed under the terms of the Creative Commons Attribution 4.0 International Public License (CC-BY 4.0), a copy of which is available at: (<https://creativecommons.org/licenses/by/4.0/legalcode>). This license permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.